



Bartovation LLC.  
2485 47<sup>th</sup> St.  
Astoria, NY 11103

## CERTIFICATION OF CONFORMANCE

Product Code: PCL03V100

Product Description: Chlorine Test Strips 0-200ppm

Expiration Date: 2 Years from date of manufacture

Each Lot of test strips is tested both for precision and accuracy.

### Precision Testing

- 10 strips from each lot are tested simultaneously in a 25 ppm chlorine standard which has been verified using a secondary test kit method, • Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

### Accuracy Testing

- Chlorine standards are prepared by serial dilution of a Chlorine stock solution.
- Standards are prepared at values of 25, 50, 100, and 200 ppm.
- Each standard is verified using a secondary test kit method.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart, • Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- % color unit.

### Shelf-Life

- An expiration date for each lot is provided. The minimum shelf-life is 2 years when the strips are stored properly in a cool, dry environment out of direct sunlight.

The lot of Chlorine strips listed above has been tested and found to meet established specifications. This lot has been released for sale.

\_\_\_\_\_ 10/17/19

Scott Bartman  
Owner | Bartovation LLC.