



Bartovation LLC.
2483 47th St.
Astoria NY 11103
contact@bartovation.com

Certificate of Conformance

Product Code: PCL04V100
Product Description: Chlorine Test Strips 0-300 ppm
Lot Number: As Stated | Expiration Date: As Stated

Each Lot of test strips is tested both for precision and accuracy.

Precision Testing

- 10 strips from each lot are tested simultaneously in a 25 ppm chlorine standard which has been verified using a secondary test kit method.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

Accuracy Testing

- Chlorine standards are prepared by serial dilution of a Chlorine stock solution.
- Standards are prepared at values of 25, 50, 100, 200, and 300 ppm.
- Each standard is verified using a secondary test kit method.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- ½ color unit.

Shelf Life

An expiration date for each lot is provided. The minimum shelf-life is 2 years when the strips are stored properly in a cool, dry environment out of direct sunlight.

The lot of Chlorine strips listed above has been tested and found to meet established specifications. This lot has been released for sale.

Scott Bartman 8/3/2025

Scott Bartman | Owner | Bartovation LLC.