



Bartovation LLC.
2483 47th St.
Astoria, NY 11103

CERTIFICATION OF CONFORMANCE

Product Code: PCL05V100

Product Description: Chlorine Test Strips, 0-1000 ppm

Lot Number: As Stated

Expiration Date: As Stated

Each Lot of test strips is tested both for precision and accuracy.

Precision Testing

- 10 strips from each lot are tested simultaneously in a chlorine standard which has been verified using a secondary test kit method.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

Accuracy Testing

- Chlorine standards are prepared by serial dilution of a chlorine stock solution.
 - Standards are prepared at values of 50, 100, 250, 500 and 1000 ppm.
 - Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- 1/2 color unit.

<u>Standard Value (ppm)</u>	<u>Acceptable Range (+/- % color unit)</u>
1000	750-1000+
500	375-750
250	175-375
100	75-175
50	25-75
0	0-25

This lot was tested and found to meet specifications. It is released for sale.

Scott Bartman 10/1/2022

Scott Bartman

Owner| Bartovation LLC.