



Bartovation LLC.  
2483 47th St  
Astoria, NY 11103

## CERTIFICATION OF CONFORMANCE

Product Code: PCL06V100

Product Description: Chlorine Test Strips, 0-2000 ppm

Lot: All Lots | Expiration Date: As Stated

Each Lot of test strips is tested both for precision and accuracy.

### Precision Testing

- 10 strips from each lot are tested simultaneously in a chlorine standard which has been verified using a secondary test kit method.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

### Accuracy Testing

- Chlorine standards are prepared by serial dilution of a chlorine stock solution.
- Standards are prepared at values of 25, 50, 200, 500, 800, 1500, and 2000 ppm.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- 1/2 color unit.

<u>Standard Value (ppm)</u>	<u>Acceptable Range(+/-1/2 color unit)</u>
2000	1750-2000+
1500	1150-1750
800	650-1150
500	350-650
200	125-350
50	37.5-125
25	12.5-37.5

**Scott Bartman** 5/3/2021

Scott Bartman | Bartovation LLC.