



2483 47th St.
Astoria, NY 11103
Phone: 707-255-2029
Email: contact@bartovation.com

CERTIFICATE OF CONFORMANCE

Product Part Number: PCL07V100

UPC: 649910801163

Product Title: Extra High Level Chlorine Plastic Strips, 0-10,000 ppm

Lot #: As Stated

Expiration Date: As printed on bottle

Each lot of test strips is tested for both precision and accuracy.

Precision Testing:

- 10 strips from each lot are tested simultaneously in a chlorine standard which has been verified using a secondary test kit method.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (no variation in color) for the lot to be acceptable.

Accuracy Testing:

- Chlorine standards are prepared by serial dilution of a chlorine stock solution.
- Standards are prepared at values of 1000, 2500, 5000, 7500, & 10000 ppm.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- 1/2 color unit.

<u>Standard Value (ppm)</u>	<u>Acceptable Range (+/- 1/2 color unit)</u>
10,000	8,750-10,000+
7,500	6,250-8,750
5,000	3,750-6,250
2,500	1,750-3,750
1,000	500-1,750
0	0-500

Brendan Clancy
Operations Manager
Bartovation LLC
orders@bartovation.com