



Bartovation LLC.
2483 47th St.
Astoria NY 11103
contact@bartovation.com

Certificate of Conformance

Product Code: PCL10V100

Product Description: Potassium Iodide Starch Test Papers

Lot Number: As Stated | Expiration Date: As Stated

Each Lot of test strips is tested both for precision and accuracy using chlorine standards.

Precision Testing

- 10 strips from each lot are tested simultaneously in a chlorine standard which has been verified using a secondary test kit method.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

Accuracy Testing

- Chlorine standards are prepared by serial dilution of a Chlorine stock solution.
- Standards are prepared at values of 10, 50, 100, 200 ppm.
- Each standard is verified using a secondary test kit method.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- 1/2 color unit

<u>Standard Value (ppm)</u>	<u>Acceptable Range (+/- 1/2 color unit)</u>
200	150-200+
100	75-150
50	25-75
10	0-25

Scott Bartman 1/2/2025

Scott Bartman | Owner | Bartovation LLC.