



Bartovation LLC.  
2483 47th st.  
Astoria, NY 11103

## **CERTIFICATE OF CONFORMANCE**

Product Code: PGA05V50  
Product Description: Humidity 4-Pad Test Strips  
Lot Number: As Stated  
Expiration Date: As Stated

Each lot of test strips is tested for precision and accuracy.

### Precision Testing

- 10 strips from each lot are tested for response to a 75%RH standard.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

### Accuracy Testing

- Humidity standards are prepared at %RH values of 25, 50, 75, and 100 using saturated salt solutions.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- ½ color unit.

### Shelf-Life

- An expiration date for each lot is provided. The minimum shelf-life is 2 years when the strips are stored properly in a cool, dry environment out of direct sunlight.

The lot listed above has been tested and found to meet established specifications. This lot has been released for sale.

***Scott Bartman*** Date: 3/14/2025

Scott Bartman

Owner | Bartovation LLC.