

Certificate of Conformance

Product Code: PQT01V100

Product Description: Low Level QAC Test Strips

Expiration Date: 2 Years from date of manufacture

Each Lot of test strips is tested both for precision and accuracy.

Precision Testing

- 10 strips from each lot are tested simultaneously in a QAC standard at 50 ppm which has been verified using a secondary test method.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

Accuracy Testing

- QAC standards are prepared by serial dilution of a QAC stock solution.
- Standards are prepared at values of 10, 50, 100 ppm.
- Each standard is verified using a secondary test method.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart. • Test strips are accepted if the color developed for the standard being evaluated visually falls within $\pm 1/2$ color unit.

<u>Standard Value (ppm)</u>	<u>Acceptable Range ($\pm 1/2$ color unit)</u>
100	75-100+
50	35-75
10	7.5-15

 10/2/19

Scott Bartman

Owner | Bartovation LLC.