



Bartovation LLC.
2483 47th St.
Astoria NY 11103
contact@bartovation.com

Certificate of Conformance

Part Number: PSP01V100

Product Description: Lead Acetate Test Papers

Lot Number: As Stated | Expiration Date: As Stated

Prior to release for sale all lots of Lead Acetate Test Papers are tested in accordance to established quality standards. Lead Acetate Test Strips are evaluated for a positive reaction to sulfide standards at 250 ppm and 5ppm. To be released for sale, a positive result is required for both standards.

This product was tested and found to meet specifications.

Scott Bartman Date: 10/14/2024

Owner | Bartovation LLC.