



Bartovation LLC.
2483 47th St.
Astoria NY 11103
contact@bartovation.com

Certificate of Conformance

Product Code: PSS01V25

Product Description: Ammonia Test Strips 0-100 ppm [Vial of 25 Strips]

Lot Number: As Stated | Expiration Date: As Stated

Each Lot of test strips is tested both for precision and accuracy.

Precision Testing

- 10 strips from each lot are tested simultaneously in a 10 ppm ammonia standard
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

Accuracy Testing

- Ammonia standards (as Ammonia-Nitrate) are prepared at values of 0, 10, 25, 50 and 100 ppm.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- ½ color unit.

Shelf-Life

- An expiration date for each lot is provided. The minimum shelf-life is 2 years when the strips are stored properly in a cool, dry environment out of direct sunlight.

The test strips listed above have been tested and found to meet established specifications and have been released for sale.

Scott Bartman Date: 1/4/2024

Owner | Bartovation LLC.