



Bartovation LLC.
2483 47th St
Astoria, NY 11103
707-255-2029
contact@bartovation.com

CERTIFICATE OF CONFORMANCE

Product Code: PSS02V50

Product Description: Ascorbic Acid Test Strips

Lot Number: As Stated | Expiration Date: As Stated

Each Lot of test strips is tested both for precision and accuracy.

Precision Testing

- 10 strips from each lot are tested for response to an Ascorbic Acid standard.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

Accuracy Testing

- Ascorbic Acid standards are prepared at 0, 0.01%, 0.05%, 0.1%, and 1%.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- ½ color unit.

Shelf-Life

An expiration date for each lot is provided. The minimum shelf-life is 2 years when the strips are stored properly in a cool, dry environment out of direct sunlight.

The lot of ASC strips listed above has been tested and found to meet established specifications. This lot has been released for sale.

Scott Bartman

12/1/2020

Scott Bartman
Owner | Bartovation LLC.