



Bartovation LLC.
2483 47th St.
Astoria NY 11103
contact@bartovation.com

CERTIFICATE OF CONFORMANCE

Product Code: PWQ05V50

Product Description: Iron Test Strips 0-100 ppm

Lot Number: As Stated | Expiration Date: As Stated

Each Lot of test strips is tested both for precision and accuracy.

Precision Testing

- 10 strips from each lot are tested for response to an Iron standard solution.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

Accuracy Testing

- Iron standards are prepared at 0, 2, 5, 10, 25, 50, and 100 ppm.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- ½ color unit.

Shelf-Life

An expiration date for each lot is provided. The minimum shelf-life is 2 years from the date of manufacture when the strips are stored properly in a cool, dry environment out of direct sunlight.

All lots of strips listed above have been tested and found to meet established specifications.
This lot has been released for sale.

Scott Bartman **Date: 1/3/23**

Scott Bartman | Owner | Bartovation LLC.