



Bartovation LLC.
2483 47th St.
Astoria NY 11103
contact@bartovation.com

Certificate of Conformance

Produce Code: PWQ09V50

Product Description: : Phosphate (Phosphorus) Test Strips

Lot Number: As Stated Expiry Date: As Stated

Each Lot of test strips is tested both for precision and accuracy.

Precision Testing

- 10 strips from each lot are tested for response to a Phosphate standard.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

Accuracy Testing

- Phosphate standards (as Phosphorus) are prepared at 0, 10, 20, 50, and 100 ppm.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against the color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- ½ color unit.

Shelf-Life

- An expiration date for each lot is provided. The minimum shelf-life is 2 years when the strips are stored properly in a cool, dry environment out of direct sunlight.

This lot of phosphate test strips listed above has been tested and found to meet established specifications. This lot has been released for sale.

Scott Bartman Date: 08/19/2022

Scott Bartman

Owner | Bartovation LLC.