



2483 47th St.  
Astoria, NY 11103  
Phone: 707-255-2029  
Email: [contact@bartovation.com](mailto:contact@bartovation.com)

## CERTIFICATE OF CONFORMANCE

**Product Part Number:** PGA03S24

**Product Title:** Chlor Phenol Red (CPR) 8 in x 10 in Test Paper Sheets

**Lot #:** All Lots

These test sheets have been immersed in solution containing indicator grade 0.15% Chlorophenol Red (CAS# 4430-20-0).

Each sheet is sensitive to pH. The sheets will turn pink/purple/violet when exposed to pH levels above 7.0. Higher ionic strength and/or higher pH samples will display a greater color intensity. The test papers are also reactive to certain vapors (example: ammonia).

Each batch of test paper is tested before release for sale using tap water and pH 9 buffer solutions. A positive color change is required before the paper is approved for sale.

This lot was tested and found to meet specifications. It is released for sale.

*Scott Bartman 1/24/202*

Scott Bartman

Owner

Bartovation LLC.