



Bartovation LLC.
2483 47th St.
Astoria, NY 11103

CERTIFICATION OF ANALYSIS

Product Code: PPH04B50

Product Description: pH 3-6 Short Range Test Strips [Bag of 50 Strips]

Lot Number: All lots

Expiration Date: Two years from manufacture date

Each Lot of test strips is tested both for precision and accuracy.

Precision Testing

- 10 strips from each lot are tested simultaneously in a pH standard which has been verified using a secondary test kit method.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

Accuracy Testing

- pH standards are prepared by dilution of pH buffer stock solution.
- Standards are prepared at values of 3.0, 3.5, 4.0, 4.5, 5.0, 5.5, 6.0.
- Each standard is verified using a secondary test kit method.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- ½ color unit.

Scott Bartman 12/1/2025

Scott Bartman

Owner | Bartovation LLC.