



Bartovation LLC.
2483 47th St.
Astoria NY 11103
contact@bartovation.com

CERTIFICATE OF CONFORMANCE

Product Code: PPX05V100

Product Description: Peroxide Test Strips 0-25ppm

Lot Number: As Stated

Expiration Date: As Stated

The peroxide test strips consist of a chemically treated cellulose test pad (5mm²) adhered to a vinyl plastic backing. The paper is treated with a chemical mixture consisting of an enzyme specific for peroxide, an indicator dye suitable for the range required, and various other chemicals – buffers, surfactants, etc.

Prior to release for sale, each lot of test strips is tested both for precision and accuracy. A verified color chart is used to evaluate the lot.

Precision Testing

- 10 strips from each lot are tested simultaneously in a peroxide standard which has been verified using a secondary test kit method.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

Accuracy Testing

- Peroxide standards are prepared by serial dilution of ACS grade 30% peroxide.
- Standards are prepared at values of 0.5, 2, 5, 10, and 25 ppm.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- ½ color unit.

<u>Standard Value (ppm)</u>	<u>Acceptable Range (+/- ½ color unit)</u>
25	12.5-37.5
10	5-15
5	2.5-7.5
2	1-3
0.5	.25-.75
0	0-.25

Scott Bartman 4/10/26

Scott Bartman
Owner | Bartovation LLC.