

Certificate of Conformance

Product Code: PWQ01V25

Product Description: Ammonia 0-6ppm Aquarium Test Strips [Vial of 25 Plastic Strips]

Lot Number: As Stated | Expiration Date: As Stated

Each lot of test strips is tested both for precision and accuracy.

Precision Testing

- 10 strips from each lot are tested simultaneously in a 1.0 ppm ammonia standard
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

Accuracy Testing

- Ammonia standards (as Ammonia-Nitrate) are prepared at values of 0, 0.5, 1.0, 3.0 and 6.0 ppm.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- ½ color unit.

Shelf-Life

An expiration date for each lot is provided. The minimum shelf-life is 2 years when the strips are stored properly in a cool, dry environment out of direct sunlight.

Scott Bartman Date: 6/17/2026

Scott Bartman | Owner | Bartovation LLC.