



Bartovation LLC.
2483 47th St.
Astoria NY 11103
contact@bartovation.com

CERTIFICATE OF CONFORMANCE

Product Code: PPA04V50

Product Description: Extra High Level Peracetic Acid Test Strips, 0-3000ppm

Lot Number: As Stated | Expiration date: As Stated

The PAA test strips consist of a chemically treated cellulose test pad (5mm²) adhered to a vinyl plastic backing. The paper is treated with a chemical mixture consisting of an indicator dye suitable for the range required, and various other chemicals – buffers, surfactants, etc.

Prior to release for sale, each lot of test strips is tested both for precision and accuracy. A verified color chart is used to evaluate the lot.

Precision Testing

- 10 strips from each lot are tested simultaneously in a PAA standard which has been verified using a secondary test kit method
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable

Accuracy Testing

- PAA standards are prepared by serial dilution from a commercially available PAA concentrate.
- Standards are prepared at values of 0, 500, 1000, 1500, 2300 and 3000 ppm.
- Each standard is verified using a secondary test kit method.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- ½ color unit.

Expiration Date

The test strips are guaranteed to perform to specification until the expiration date printed on the bottle assuming the test strips are stored in the original container and kept in a cool, dry environment out of direct sunlight.

This lot was tested and found to meet specifications. It is released for sale.

Scott Bartman Date: 7/7/2026

Scott Bartman

Owner | Bartovation LLC.