Bartovation LLC.
2483 47th St.
Astoria, NY 11103

CERTIFICATE OF CONFORMANCE

Product Code: PQT01V100
Product Description: Low Level QAC Test Strips, 0-100ppm

Lot Number: All Lots | Expiration Date: As Stated

Each lot of test strips is tested both for precision and accuracy.

Precision Testing

- 10 strips from each lot are tested simultaneously in a QAC standard at 50 ppm which has been verified using a secondary test method.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

Accuracy Testing

- QAC standards are prepared by serial dilution of a QAC stock solution.
- Standards are prepared at values of 10, 50, 100 ppm.
- Each standard is verified using a secondary test method.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- ½ color unit.

Standard Value (ppm)	Acceptable Range (+/- ½ color unit)
100	75-100+
50	40-75
30	25-40
20	15-25
<10	0-15

Scott Bartman 11/15/2026

Owner | Bartovation LLC.